Infrared Heat Lamp

Instruction Manual

Assembly Instructions: Connecting the Casters and Pole

Components

- Control panel with height adjustment knob
- Lower pole with safety cover
- 5-leg caster base

Assembly Steps

1. **Prepare the pole**
   - Remove the rubber tip from the metal end of the pole before installation
2. **Insert the pole**
   - Place the pole into the center hole of the 5-leg caster base
   - Ensure the pole is properly aligned
3. **Secure the connection**
   - Turn the pole clockwise to lock it firmly into position
   - Verify the connection is stable before use

Operating Instructions

I. Preparation

Before using the device:

- Read all operation instructions thoroughly
- Place the unit on a flat, stable surface
- Wear appropriate light clothing for treatment

II. Operation Procedure

1. Power Setup

- Connect the power cord to a 110V outlet

2. Initial Settings

- Set the intensity to 0 on the main control panel
- Turn the timer dial clockwise to set treatment duration

3. Temperature Adjustment

- Turn the intensity dial on the main control panel clockwise to set desired temperature
- Position the lamp head approximately 18 inches from the treatment area
- Adjust distance and temperature as needed

4. Remote Control Operation (if applicable)

- Plug in the cord and locate the timer on the lamp
- Turn the timer counter-clockwise until you hear a click for ON setting
- The heat can be adjusted using the remote controller
- If the remote controller is not functioning, the intensity dial on the main control panel is disabled

5. Treatment Guidelines

- Recommended treatment time: 30 minutes per session
- Avoid overheating
- Allow the unit to cool down before resuming if using an optional plastic screen that may melt if overheated

6. Post-Treatment

- Unplug the unit after treatment

Safety Precautions

Medical Contraindications

⚠ **Warning:** Patients with the following conditions must consult a physician before using this product:

- Acute pain
- Malignant tumors
- High fever
- Heart impairment
- Pregnancy

During Operation

Monitor the following during use:

1. Verify the lamp operates properly
2. Ensure each cord is securely connected
3. Confirm the lamp head is safely positioned above the treatment area
4. Do not exceed recommended treatment time of 30 minutes per session
5. If the lamp or patient shows any anomaly, stop operation immediately and take proper measures
6. Operation must be performed by a well-trained practitioner
7. Turn off power after completing treatment

Storage and Maintenance

After use, please observe the following:

1. Turn off the power
2. Do not pull on the cord or exert excessive force on connective parts
3. Store the main unit and auxiliary parts in a clean, dry place
4. Avoid direct sunlight and high concentrations of sodium ions in the atmosphere
5. Keep out of reach of children

Service and Warranty Information

Service Policy

- Do not return defective parts without proper written authorization
- All transportation costs and shipping damage incurred while submitting parts for repair or replacement are the responsibility of the purchaser

Important: Failure to follow the instructions and warnings may result in personal injury or damage to the product and will void the product warranty.

Virtual Health Options LLC
Website: www.virtualhealthoptions.com
Phone: 619-825-6884

MADE IN KOREA